

PATENT APPLICATION BASED ON:

Docket No:

82295/WFN

Inventors:

Xiaohui Wang

David H. Foos

Attorney:

William F. Noval

**METHOD FOR ACQUIRING A RADIATION IMAGE OF A LONG BODY**  
**PART USING DIRECT DIGITAL X-RAY DETECTORS**

Commissioner for Patents

Attn: Box Patent Application

Washington, DC 20231

Express Mail Label No: EL656964126US  
Date: April 30, 2001

094558-043001

**METHOD FOR ACQUIRING A RADIATION IMAGE OF A LONG BODY**  
**PART USING DIRECT DIGITAL X-RAY DETECTORS**

**FIELD OF THE INVENTION**

5                   This invention relates in general to digital radiography, and in particular to the imaging of a long human body part, such as the spine or the legs, using a direct digital x-ray detector system.

**BACKGROUND OF THE INVENTION**

10                   When a long segment of the human body is imaged using a conventional screen-film technique, special cassettes and films of extended length are used, such as 30 x 90 cm and 35 x 105 cm. As shown in Fig. 1, both the x-ray tube 101 and the cassette 103 are kept stationary during the exam, and an image of the patient is acquired in a single exposure. As medical institutions are migrating from analog screen-film systems to digital modalities, such as flat-panel detector  
15                   based digital radiography (DR) systems, this type of exam imposes a significant challenge. This is because the sizes of digital detectors are limited by cost and technology. The largest flat panel DR plates available today are limited to 43 x 43 cm. A detector of this size is capable of imaging only a portion of the body part at a time and thus is inadequate for performing imaging exams of longer length body  
20                   parts such as the full spine or full leg. To address this problem, multiple exposures at varying patient positions must be acquired with the assumption that the patient can keep still during the exam. The individual images are then stitched together to reconstruct a larger composite image. Fig. 2A shows this technique (see: U.S. Patents 5,123,056 and 4,613,983). First, the patient 200 is exposed at a  
25                   first position defined by the x-ray tube position 201 and the detector position 203. The collimator of the x-ray tube is adjusted such that the x-rays 202 can barely cover the detector while protecting the patient from unnecessary radiation in the non-imaging related regions. Second, after both the x-ray tube and the detector are translated parallel along axis 210 and 211 to a second position, as indicated by  
30                   206 and 208, a second exposure of the patient is conducted. There can be a slight overlap between the consecutive detector coverage in order to facilitate image

09045588.043001  
T00E10.8854860

stitching. This process can be continued until the full length of the body part to be examined is imaged. However, this stereovision-like image acquisition geometry has two major drawbacks: (1) mechanical complexity because both the detector and the x-ray tube need to be translated, and (2) inherent geometric distortion that makes precise image stitching almost impossible. As shown in Fig. 2B, if an object AB is situated in both the exposure coverage of x-ray tube position 201 and 206, it will be imaged at the corresponding detector position as  $AB_0$  and  $AB_1$ , respectively, due to the divergence of the x-rays. With  $AB_0$  pointing downward but  $AB_1$  pointing upward, the same object has apparently created two different images. Evidently,  $AB_0$  and  $AB_1$  can not be registered together in the stitched image. Therefore, theoretically the images acquired at different tube positions can never be seamlessly and precisely stitched. This severity of this problem becomes worse if the body part get thicker. There is a need to develop an imaging method with DR that not only is simple in design but also can provide distortion-free images for stitching.

### SUMMARY OF THE INVENTION

According to the present invention, there is provided a solution to the problems and fulfillment of these needs.

According to a feature of the present invention, there is provided a method for acquiring an elongated radiographic image comprising: positioning an elongated object between a source of x-rays and a digital image capture device having a known imaging dimension which is less than a like dimension of said elongated object; moving said device in a direction parallel to said known imaging dimension to sequential contiguous positions to acquire a sequence of radiographic images of said elongated object; and rotating said source of x-rays about an axis perpendicular to said direction of moving said device in coordination with said moving project said x-rays from said source toward said device.

0945588-043001

sub  
A1

## ADVANTAGEOUS EFFECT OF THE INVENTION

The invention has the following advantages.

1. Eliminates the stereovision-like geometric distortion caused by the current imaging method with DR systems, and allows precise registration of the partial images to reconstruct a larger composite image.

## BRIEF DESCRIPTION OF THE DRAWINGS

Fig. 1 is a diagrammatic view showing an imaging method used with screen-film systems.

Figs. 2A and 2B are diagrammatic views showing an imaging method used with DR systems and showing the distortion introduced by the stereovision-like effect from the two subsequent exposures.

Figs. 3A and 3B are diagrammatic views illustrating a method of the invention that allows multiple partial images to be taken with similar acquisition geometry to that obtained using screen-film systems.

## DETAILED DESCRIPTION OF THE INVENTION

An overview illustration of the present invention is shown in Figs. 3A and 3B.

The digital radiographic image capture device or detector 303 can translate along axis 311 freely to various positions, the x-ray tube 301 does not translate but it tracks the detector movement by rotation along a fixed axis 310 and always directs its exposure toward the detector 303. First, the patient 300 is exposed at a first position defined by the detector position A and the x-ray tube orientation  $T_1$ . The collimator of the x-ray tube 301 is adjusted such that the x-rays 302 just cover the detector 303 for imaging. Second, after the detector 301 is translated to a second position, as indicated by B, the tube 301 rotates to a new orientation, as indicated by  $T_2$ , and the aperture of the collimator is adjusted again if necessary, a second exposure of the patient is conducted. This process can be continued until the full length of the body part to be examined is imaged. There can be a slight overlap between the consecutive detector position coverage in

15                Although the present invention has been described with respect to acquiring radiographic images of elongated body parts, it will be understood that any elongated object may be subject of the radiographic imaging techniques of the invention. Digital radiographic image capture device may be any such device but preferably is a flat panel radiographic detector known to those skilled in the art.

20                   The invention has been described in detail with particular reference  
to certain preferred embodiments thereof, but it will be understood that variations  
and modifications can be effected within the spirit and scope of the invention.

**PARTS LIST**

- 100 patient
- 101 x-ray tube and collimator
- 102 -ray coverage
- 103 screen-film
- 200 patient
- 201 x-ray tube and collimator at a first location
- 202 x-ray coverage at a first location
- 203 DR detector at a first location
- 206 x-ray tube and collimator at a second location
- 207 x-ray coverage at a second location
- 208 DR detector at a second location
- 210 moving axis of x-ray tube and collimator from a first location to a second location
- 211 moving axis of DR detector from a first location to a second location
- 300 patient
- 301 x-ray tube
- 302 x-ray coverage at a first detector location
- 303 DR detector
- 305 x-ray coverage at a second detector location
- 310 x-ray tube rotation axis
- 311 moving axis of DR detector from a first location to a second location
- 320 image processor

09845588.043001